

CBSPD Exam Prep Questions and Answers

2023 - 2024 (Verified Answers)

1. Semi critical devices should be processed by which disinfection method?-

ANS High-level disinfection (at least)

2. Critical devices should be processed by which disinfection method?

ANS Sterilization

learnexams

3. Non critical devices should be processed by which disinfection method? **ANS**

Sanitization

4. Name the three different types of endoscopic scopes **ANS** Rigid, flexible, and semi-rigid

5. According to Spaulding's classification of medical devices, a semi critical item requires which form of disinfection? **ANS** High-level disinfection

6. In a non-tabletop sterilizer, a gauge on the main steam line, usually located at the back of the sterilizer, tells you the steam pressure **ANS** As it enters the jacket

7. One reason for adding chemicals to the boiler feed water is to **ANS** Minimize contaminates

8. The Bowie-dick is considered which type of monitoring process? **ANS** Chemi- cal Monitoring

learnexams

9. Routine biological testing of steam sterilizers should be performed no less than **ANS** Weekly

10. All loads containing implantable devices should be **ANS** Monitored with a bio- logical indicator

11. Unless otherwise directed by the sterilizer manufacturer, the ethylene oxide process challenge device (bi test pack) is placed in the **ANS** Geometric center of the load

12. Which of the following is compatible with Hydrogen peroxide ozone sterilization? **ANS** Stainless steel instruments

13. Biological monitoring of the hydrogen peroxide-ozone sterilization process should be performed **ANS** At least daily

14. "The removal of contamination from an item to the extent necessary for further processing or for its intended use" is the Association for the Advancement of Medical Instruments definition of which of the following **ANS** - Cleaning

learnexams

15. Negative pressure refers to **ANS** Air exhausted directly to the outside

16. When sorting items in the decontamination area, which of the following would require manual cleaning? **ANS** Powered Equipment